

# CERTIFICATE OF CALIBRATION

Issued 2025-09-20

CERTIFICATE NO  
**DEMO-005-2**

## Lakeside Machining Co.

### RECORD INFORMATION

Control No DEMO-005  
Serial No TT1507523

### INSTRUMENT INFORMATION

Type	Torque Wrench	Range or size	10–50 N·m
Manufacturer	Norbar	Location	Assembly
Model	TTi 50		

### CALIBRATION INFORMATION

Date calibrated	2025-09-20
Next calibration due	<b>2026-09-18</b>
Calibration interval	12 months
Calibrated by	In-house

### RESULTS

**As found: FAIL** **As left: PASS** **Fit for purpose: YES**

Adjustments made: Yes · Repairs made: No

### MEASUREMENT DATA

#	PARAMETER	NOMINAL	TOL ±	AS FOUND	AS LEFT	UNIT	RESULT
1	20 N·m CW	20	0.8	21.4	20.1	N·m	<b>PASS</b>
2	35 N·m CW	35	1.4	36.9	35.2	N·m	<b>PASS</b>
3	50 N·m CW	50	2	53.1	50.3	N·m	<b>PASS</b>

### STATEMENT

The instrument identified on this certificate was calibrated using the procedure and reference standards stated herein. Results relate only to the item calibrated and are certified accurate on the date of calibration by the identified technician.

### REVIEW AND APPROVAL

Calibrated by  
**In-house**  
2025-09-20

Approved by